

ABSOLUTE BEAM EMITTANCE MEASUREMENTS AT RHIC USING IONIZATION PROFILE MONITORS*

M. Minty[#], R. Connolly, C. Liu, T. Summers, S. Tepikian, BNL, Upton, NY, USA

Abstract

In this report we present studies of and measurements from the RHIC ionization profile monitors (IPMs). Improved accuracy in the emittance measurements has been achieved by (1) continual design enhancements over the years, (2) application of channel-by-channel offset corrections and gain calibrations in the beam profile measurements and (3) use of measured beta functions at the locations of the IPMs. The removal of systematic errors in the emittance measurements was confirmed by the convergence of all four planes of measurement (horizontal and vertical planes of both the Blue and Yellow beams) to a common value during beam operation with stochastic cooling. Consistency with independent measurements (luminosity-based using zero degree counters) at the colliding beam experiments STAR and PHENIX was demonstrated.

INTRODUCTION

In the past, comparisons between emittance measurements from ionization profile monitors, Vernier scans (using as input measured rates from zero degree counters, ZDCs), the polarimeters and the Schottky detectors evidenced significant variations of up to 100%. In this report we present studies of the RHIC ionization profile monitors (IPMs). After identifying and correcting for systematic instrumental errors in the beam size measurements, we present experimental results showing that the remaining dominant error in beam emittance measurements was imprecise knowledge of the local beta functions. After application of measured beta functions, thanks to full 3-D stochastic cooling we demonstrate that precise measurements of the absolute emittances result. Also, consistency between the emittances measured by the IPMs and the ZDCs was demonstrated.

BRIEF HISTORY OF THE RHIC IPMS

The design of the RHIC IPMs has evolved over time with continuous improvements. The first prototype was built and tested in 1996 [1] with first measurements in RHIC in 1999 [2, 3]. In 2002 two changes were made motivated by experiences with beam: shielding was added upstream of the detectors to prevent signal contributions from upstream beam losses and the electrodes were made longer to avoid electron clouds from migrating into the region of the detector [4]. In 2005 fast signal gating was added to avoid depletion of the multichannel plate (MCP)

detector and better isolation of the detector from the electromagnetic fields of the beam was implemented [4]. As the beam intensities increased, this latter effect was further suppressed with a new design in 2007 which placed all electronics inside a Faraday cage outside of the path of the beam's image current [5]. The prototype for this new design [5] was implemented for the Yellow Ring vertical plane (YV) in 2008. In 2010 the new design [5] was implemented for both the Blue and Yellow Ring horizontal planes (BH, YH) and these IPMs were moved to locations where the beam sizes were larger. The new IPM design [5] was implemented in the last remaining plane, Blue Ring vertical (BV), in 2013.

INSTRUMENTAL SYSTEMATIC ERRORS

During the FY11 RHIC run [6], it was found that when the beams were brought into collision at a new third experiment (AnDY), the vertical beam size measurements from the IPMs changed considerably. To investigate further, measurements were taken while scanning the beam across the area of the detector. The measurements revealed significant damage to the multi-channel plate detectors (MCP) due to depletion [7] however with corrections applied, the position sensitivity still remained (since the beam sizes at full energy were small compared to the region of the MCP depletion). Other issues concerned variations between measurements, which were not small compared to expectation based on the statistical properties (i.e. ionization cross sections) of the measurements, and channel-to-channel variations within a single measurement.

CHANNEL-BY-CHANNEL OFFSET CORRECTION AND GAIN CALIBRATIONS

A conceptual view of the IPM and photographs are shown in Fig. 1. The signals from the MCP (bottom left) are processed through 64 channels as seen on the anode board (bottom right) and transferred to amplifiers through 64 ceramic-beaded wires (bottom left). Using the previously acquired data (beam size measurements as a function of beam centroid position), the channel offsets were determined by measurements with beam passing across the MCPs but not above the specific channels of interest. After applying the offset corrections to all the measured profiles and removing bad channels, the gain calibrations were obtained using the following automated procedure. For each IPM:

- (1) Fit each of the (20 to 30 or so) profiles in the calibration scan with a Gaussian and compute the chi-squared, χ^2 .

* Work supported by Brookhaven Science Associates, LLC under Contract No. DE-AC02-98CH10886 with the U.S. Department of Energy
minty@bnl.gov

- (2) Calculate a figure of merit equal to the mean chi-squared $\langle \chi^2 \rangle$.
- (3) For a given channel (the RHIC IPMs have 64 readout channels), scale the channel gain and repeat steps 1-2.
- (4) Iterate steps 1-3 over a range of channel gain scale factors.
- (5) Perform a polynomial fit to the resultant $\langle \chi^2 \rangle$ versus scaled channel gain. Examples are shown in Fig. 2.
- (6) Implement the channel gain so found (with minimum $\langle \chi^2 \rangle$).
- (7) Repeat steps 1-6 for all 64 channels.

Figure 3 shows IPM profile measurements from the FY11 RHIC run [5] without channel-by-channel corrections (top) and from the FY12 RHIC run with corrections (bottom). The new and simplified header display now includes a fit error which until these corrections were applied was not useful in evaluation of the quality of the profile measurements. As can be seen, the Gaussian fit algorithm was considerably better constrained after implementing the channel-by-channel offset corrections and gain calibrations.




Figure 1: Conceptual view (top) and photographs (bottom) of new RHIC IPM [5]. The MCP (left) dimensions are 8 cm by 10 cm. The signals from the MCP are collected by the 64-channel anode board (right) and read out through the ceramic-beaded wires (left).




Figure 2: Examples of polynomial fit to figure of merit, $\langle \chi^2 \rangle$, versus channel gain scaling factor for two different IPM channels.




Figure 3: Example beam profile measurements before (top) and after (bottom) implementation of IPM offset and gain corrections. In each subplot, shown in the top row are the Blue ring horizontal (left) and vertical (right) profiles and in the bottom row the Yellow ring horizontal (left) and vertical (right) profiles.

EMITTANCE MEASUREMENTS USING MEASURED BETA FUNCTIONS

After removal of the systematic measurement errors in the measured beam sizes, the presence of another significant systematic error became apparent (e.g. unphysical shrinking emittances without accompanying beam loss during acceleration in FY12 and FY13 and during the “rotator ramp” with a special E-Lens optic in FY13). Since the emittance ϵ is given by the ratio of the square of the rms beam size σ_β to the beta function β at the location of the IPM, the value assumed for the beta function was suspect. As an illustration, the reported horizontal and vertical emittances of both the Blue and Yellow beams did not converge to equal values which with stochastic cooling [8] is expected to result [8, 9]. This was the case in the FY11 RHIC Run when stochastic cooling was applied in the vertical planes with damping of both transverse planes achieved through betatron




Figure 4: Evolution of the transverse beam emittances with 3D stochastic cooling [9] during the FY14 RHIC Run derived from the IPM profile measurements in all 4 planes (horizontal and vertical in the Blue and Yellow Rings) during two physics stores derived using model beta functions (top) and, for the same two stores, using beta functions interpolated from measurements from nearby beam position monitors (bottom).

coupling [8]. Similarly, in FY14 with full 3-D stochastic cooling [9], the emittances did not converge to identical values as shown in Fig. 5 (upper plot).

During the FY13 and FY14 RHIC Runs, the beta functions measured at the beam position monitors bracketing the IPMs were used to interpolate to the positions of the IPMs [10]. A comparison between the model beta functions (used to date for derivation of the emittances from the beam profile measurements) and the measured beta functions obtained early in the FY14 Run is given in Table 1. The measured beta functions were used to correct the measurements (by scaling by the ratio $\beta_{\text{model}}/\beta_{\text{meas}}$). Shown in Fig. 4 (bottom) are the emittances obtained using the measured beta functions rather than the model beta functions (which were used in Fig. 4 (top)). The two physics stores shown here correspond to a time late in the FY14 Run at which time RHIC operations was fully optimized with only occasional adjustments over the course of a physics store to stochastic cooling. The adjustments were made by Operations based on observed relative increased emittances, which was indicative of over-cooling of one beam relative to the other (examples of this are visible in Fig. 4 at ~01:30 and ~10:00). The convergence of the measurements in all 4 planes to a

common value gives good confidence that the emittances are now accurately determined.

Table 1: Model and Measured Beta Functions at Store Energy During the FY14 Au+Au RHIC Run

	Blue Horizontal	Blue Vertical	Yellow Horizontal	Yellow Vertical
β_{model} (m)	202	118	206	112
β_{meas} (m)	262	109	245	174
$\beta_{\text{model}}/\beta_{\text{meas}}$	0.77	1.08	0.84	0.64

COMPARISON OF EMITTANCE MEASUREMENTS FROM THE RHIC IPMS AND FROM THE RHIC ZDCs

As mentioned in the introduction, emittance comparisons using different methods have in the past shown disparities. Here we compare the emittance measurements from the IPMs to those inferred using the experiment's ZDC counters [11] as input and analyzed in the application StoreAnalysis [12], which among other things computes the effective beam emittance taking into account the measured ZDC rates, the model beta

Beam Profile Monitors
Monday poster session




Figure 5: Comparison of rms emittance measurements during Au+Au, 100 GeV operations from the RHIC IPMs and from the experimental ZDCs (red and black lines) from the FY11 Run (Fig. 5a, no channel-by-channel offset corrections and gain calibrations), from the FY14 Run (Fig. 5b, with channel-by-channel corrections and model beta functions at the IPMs), the same data from the FY14 Run (Fig. 5c, with measured beta functions at the IPMs only) and the same data from the FY14 Run (Fig. 5d, with measured beta functions at the IPMs and at STAR and PHENIX). The horizontal time scale is 3.5 hours in all cases.

functions (assumed equal in all 4 planes), and the bunch length (for evaluation of the hourglass effect). Shown in Fig. 5 are the rms emittance measurements from the RHIC IPMs and from the ZDCs. In Fig. 5a (FY11 Run - Au+Au, 100 GeV), the IPM measurements did not at that time have channel-by-channel offset corrections or gain calibrations. In Fig. 5b (FY14 Run - Au+Au, 100 GeV) the channel-by-channel offset corrections were applied (with unity gain corrections as these were found not to be necessary since removal of certain unused electronics during the preceding shutdown [13]). Figure 5c shows the same data as Fig. 5b now with the measured beta functions applied at the IPMs. Figure 5d shows the same data as Fig. 5b using measured beta functions at the IPMs and at the ZDCs. For the latter, the measured beta functions were used (1) to determine the beam sizes at the interaction points and (2) for a small (<2%) correction to the geometric luminosity reduction due to finite bunch length (hour glass effect) [14]. From these results we observe that with all corrections applied the agreement

between emittance measurements from the RHIC IPMs and the ZDC counters is considerably improved

SUMMARY

The accuracy of emittance measurements using the RHIC IPMs has been greatly improved by the following: (1) continual design enhancements over the years [1-5], (2) application of channel-by-channel offset corrections and gain calibrations in the beam profile measurements and (3) use of measured beta functions at the locations of the IPMs. The removal of systematic errors in the emittance measurements was confirmed by the convergence of all four planes of measurement (horizontal and vertical planes of both the Blue and Yellow beams) to a common value during beam operations with stochastic cooling (Fig. 4). Consistency with independent measurements (based on the ZDCs) at STAR and PHENIX was shown to be within ~ 15% (Fig. 5).

Future studies will involve continued efforts towards more precise measurements of the beta function during acceleration and optics correction during acceleration [15-17] so that, together with measurements from the injectors, a better understanding of sources of emittance dilution between the injector synchrotron, the AGS, and RHIC and during acceleration in RHIC can be better characterized and localized and, eventually, corrected.

ACKNOWLEDGEMENTS

The RHIC IPM design and evolution thereof has been led by R. Connolly. The beautiful 2008 design had fabrication support by J. Fite, S. Jao, D. Kipp, T. Russo and C. Trabocchi and support from Vacuum, Controls and Instrumentation for installation and integration. S. Tepikian, author of the fitting algorithm, provided continuous enhancements to the online data processing application. T. Summers championed implementation of the online calibration measurements. C. Liu's development and automation of the beta function interpolation routine has been invaluable.

REFERENCES

- [1] R. Connolly et al, "A Prototype Ionization Profile Monitor for RHIC", PAC97, Vancouver, B.C., Canada, May 1997, p. 2152 (1997).
- [2] R. Connolly et al, "The RHIC Ionization Profile Monitor", PAC99, New York, NY, Mar 1999, p. 2144 (1999).
- [3] R. Connolly et al, "Beam Profile Measurements and Transverse Phase-Space Reconstruction on the Relativistic Heavy-Ion Collider", Nucl. Instr. and Meth. in Phys. Res. A443, pp. 215-222 (2000).
- [4] R. Connolly et al, "Residual-Gas Ionization Profile Monitors in RHIC", PAC05, Knoxville, TN, May 2005, p. 230 (2005).
- [5] R. Connolly et al, "Residual-Gas Ionization Beam Profile Monitors in RHIC", BIW10, Santa Fe, NM, May 2010, TUPSM010, p. 116 (2010).
- [6] G. Marr et al, "RHIC Performance for FY2011 Au+Au Heavy Ion Run", Proc. of IPAC2011, San Sebastian, Spain, Sept 2011, TUPZ038, p. 1894 (2011).
- [7] R. Connolly, M. Minty and S. Tepikian, "Calibration Scans on IPMs", APEX Weekly Meeting, http://www.cadops.bnl.gov/Instrumentation/InstWiki/index.php/File:APEX_IPM_presentation.pdf (2011).
- [8] M. Blaskiewicz, J.M. Brennan and F. Severino, "Operational Stochastic Cooling in the Relativistic Heavy-Ion Collider", Phys. Rev. Lett. 100, 174802 (2008).
- [9] M. Blaskiewicz, J.M. Brennan and K. Mernick, "Three-Dimensional Stochastic Cooling in the Relativistic Heavy Ion Collider", Phys. Rev. Lett. 105, 094801 (2010).
- [10] C. Liu et al, "Global Optics Correction in RHIC Based on Turn-by-Turn Data from ARTUS Tune Meter", NA-PAC13, Pasadena, CA, Sept 2013, TUPBA06, p.532 (2013).
- [11] C. Adler et al, "The RHIC Zero Degree Calorimeters", Nucl. Instr. and Meth. in Phys. Res. A 470, pp. 488-499 (2001).
- [12] W. Fischer, private communication.
- [13] C. Dawson, private communication.
- [14] M. Minty et al, "Absolute beam emittance measurements at RHIC using ionization profile monitors", BNL-105970-2014-IR, C-A/AP/522 (2014).
- [15] M. Minty et al, "Measurement of Beam Optics During Acceleration in the Relativistic Heavy Ion Collider", NA-PAC13, Pasadena, CA, TUPBA08, p. 538 (2013).
- [16] C. Liu, A. Marusic, M. Minty, "Implementation of Optics Correction on the Ramp in RHIC", NA-PAC13, Pasadena, CA, TUPBA05, p. 529 (2013).
- [17] C. Liu, A. Marusic, M. Minty, "Optics Measurement and Correction during Beam Acceleration in the Relativistic Heavy Ion Collider", BNL-106065-2014-IR, C-A/AP/528 (2014).