

ENTRY NO. 76

NAME OF MACHINE Cyclo-Graaff Injector DATE 8/1/81
INSTITUTION Triangle Universities Nuclear Laboratory
ADDRESS Duke University Durham, NC 27706 (USA)
TEL 919-684-8158 TELEX
IN CHARGE E.G. Bilpuch REPORTED BY F.O. Purser

HISTORY AND STATUS

DESIGN, date 1964 Model tests
ENG DESIGN, date 1965
CONSTRUCTION, date
FIRST BEAM, date (or goal) Factory 1966, site 1968
MAJOR ALTERATIONS Additional Harmonic Coils, Phase
Limiting Slits, Moveable Magnetic Channel
COST, ACCELERATOR Cyclotron \$360,000 FN Tandem
COST, FACILITY, total \$3,945M \$1.25 M
FUNDED BY DOE, NSF, HEW, N.C., Duke Univ.

ACCELERATOR STAFF, OPERATION AND DEVELOPMENT

SCIENTISTS 19 ENGINEERS
TECHNICIANS 4 CRAFTS 4
GRAD STUDENTS involved during year
OPERATED BY X Research staff or Operators
OPERATION 168 hr/wk. On target 140 hr/wk
TIME DISTR. in house 100 % Outside %
BUDGET, op & dev \$900,000
FUNDED BY DOE

RESEARCH STAFF, not included above

USERS, in house outside
GRAD STUDENTS involved during year
RESEARCH BUDGET, in house
FUNDED BY

MAGNET

POLE FACE, diameter (compact) 80 cm, R extraction 33.5 cm
R injection cm
GAP, min 5 cm, Field 12 kG
min 10 cm, Field 20.5 kG at 167,000
AVERAGE FIELD at R ext 16.4 kG Ampere turns
B max / < B > 1.25
NUMBER OF SECTORS { compact 3 } Spiral, max deg
{ separated }
SECTOR ANGLE (SSC) 45 deg

TRIMMING COILS

CONDUCTOR, material and type Hollow Core Copper
STORED ENERGY (cryogenic) MJ
POWER: main coils 30 max, kW; current stability 6 x 10^-5
trimming coils max, kW; current stability
WEIGHT: Fe 14 tons; coils tons
COOLING system water
ION ENERGY (bending limit) E/A = .15 q^2/a^2 MEV/amu
(focusing limit) E/A = q/a MeV/amu

ACCELERATION SYSTEM

DEES, number 2; angle 120 deg
BEAM APERTURE .2, .05 cm; DC Bias 1.0 kV
TUNED by, coarse moveable strap fine moveable plate
RF 12.8 or 25.0 MHz, stable +/- .1 KH
Orb F to MHz
HARMONICS, RF/Orb F, used Fundamental only
DEE-Gnd, max 40 kV, min gap cm
STABILITY, (pk-pk noise)/(pk RF volt) 0.002
ENERGY GAIN, max 120 kV/turn
RF PHASE, stable to +/- .5 deg
RF POWER input, max 40 kW
FREQUENCY MODULATION, rate /s
modulator, type
beam pulse, width

VACUUM SYSTEM

OPERATING PRESSURE 1 x 10^-6 Torr or mbar
PUMPS, No. Type Size
1-10 Diffusion Pump
N2 Traps

ION SOURCES

Ehlers, Penning Ion Gauge

INJECTION SYSTEM

Axial

EXTRACTION SYSTEM

Harmonic Precession, Electrostatic Deflector, Mag. Channel

FACILITIES FOR RESEARCH

SHIELDED AREA, fixed 1000 m^2, movable m^2
TARGET STATIONS 9 in 3
STATIONS served at same time, max 1
MAG SPECTROGRAPH, type
COMPUTER model DEC VAX 11/780
OTHER FACILITIES

CHARACTERISTIC BEAMS

Table with columns: PARTICLE, ENERGY (MeV) Goal, Achieved, CURRENT (pA) Internal, External. Rows for H- and D- ions.

SECONDARY

(part/s)

BEAM PROPERTIES

Table with columns: MEASURED, CONDITIONS. Rows for PULSE WIDTH, PHASE EXC., EXTRACT eff, RESOL, EMITTANCE.

OPERATING PROGRAMS, time distribution

BASIC NUCLEAR PHYSICS SOLID STATES PHYSICS
BIOMEDICAL APPLICAT ISOTOPE PRODUCTIONS
Basic Nuclear Physics 100%

REFERENCES/NOTES

PLAN VIEW OF FACILITY, COMMENTS, ETC.

Beam properties with internal phase limiting slits
Pulse width 3.2 R.F. Deg. 1.5 uA of 15 MeV H-
Res. delta E/E 0.1%
Burst Length 0.5 ns.