

ENTRY No. CU55

NAME OF MACHINE Meditron DATE June 1989

INSTITUTION Nihon Medi-Physics Co., Ltd. Takarazuka Facility

ADDRESS 4-2-1, Takatsukasa, Takarazuka, Hyogo, JAPAN 665

TEL 0797-71-4771 TELEX 5645669 NMP.J

IN CHARGE H. Tobiki REPORTED BY Y. Tanaka

HISTORY AND STATUS

DESIGN, date Model tests TCC CS-30
ENG DESIGN, date
CONSTRUCTION, date Aug, Nov, 1974
FIRST BEAM, date (or goal) Nov, 1974
MAJOR ALTERATIONS Addition of Yoke Iron
New Magnet PS
COST, ACCELERATOR
COST, FACILITY, total
FUNDED BY Nihon Medi-Physics Co., Ltd.
ACCELERATOR STAFF, OPERATION AND DEVELOPMENT
SCIENTISTS ENGINEERS 5
TECHNICIANS 7 CRAFTS
GRAD STUDENTS involved during year
OPERATED BY Research staff or Operators
OPERATION hr/wk, On target hr/wk
TIME DISTR. in house 1.00 % , Outside %
BUDGET, op & dev
FUNDED BY
RESEARCH STAFF, not included above
USERS, in house outside
GRAD STUDENTS involved during year
RESEARCH BUDGET, in house
FUNDED BY
MAGNET
POLE FACE, diameter (compact) .96. cm, R extraction .42. cm
R injection cm
GAP, min cm, Field kG
max cm, Field kG } at
AVERAGE FIELD at R ext 17.5 kG } Ampere turns
B max / < B >
NUMBER OF SECTORS { compact separated } Spiral, max .. deg
SECTOR ANGLE (ISS) deg
TRIMMING COILS
CONDUCTOR, material and type
STORED ENERGY (cryogenic) MJ
POWER: main coils .50. max, kW ; current stability 10^-5
trimming coils max, kW ; current stability
WEIGHT: Fe .20. tons ; coils tons
COOLING system
ION ENERGY (bending limit) E/A = q^2/a^2 MeV/amu
(focusing limit) E/A = q^2/a^2 MeV/amu
ACCELERATION SYSTEM
DEES, number 2 ; angle 90 deg
BEAM APERTURE cm ; DC Bias 1.5 kV
TUNED by, coarse fine
RF to MHz, stable +/-
Orb F to MHz
HARMONICS, RF/Orb F, used
DEE - Gnd, max kV, min gap cm
STABILITY, (pk-pk noise)/(pk RF volt)
ENERGY GAIN, max kV/turn
RF PHASE, stable to +/- deg
RF POWER input, max kW
FREQUENCY MODULATION, rate /s
modulator, type
beam pulse, width
VACUUM SYSTEM
OPERATING PRESSURE .5. x 10^-6 Torr or mbar
PUMPS, No, Type, Size 1. x 10" D.P.
1. x 4" D.P.

INJECTION SYSTEM

EXTRACTION SYSTEM

FACILITIES FOR RESEARCH

SHIELDED AREA, fixed 3 in m^2 ; movable m^2
TARGET STATIONS 3 in rooms
STATIONS served at same time, max
MAG SPECTROGRAPH, type
COMPUTER model
OTHER FACILITIES

CHARACTERISTIC BEAMS

Table with columns: PARTICLE, ENERGY (MeV) Goal, Achieved, CURRENT (pμA) Internal, External. Row for proton: 2.6, 2.6, 200, 50

SECONDARY

BEAM PROPERTIES

Table with columns: MEASURED, CONDITIONS. Rows for PULSE WIDTH, PHASE EXC, EXTRACT eff, RESOL ΔE/E, EMITTANCE

OPERATING PROGRAMS, time distribution

BASIC NUCLEAR PHYSICS .. SOLID STATES PHYSICS
BIOMEDICAL APPLICAT. ISOTOPE PRODUCTIONS

REFERENCES/NOTES

PLAN VIEW OF FACILITY, NOTEWORTHY FEATURES, COMMENTS