

ENTRY NO. CU98

NAME OF MACHINE FMI CYCLOTRON
INSTITUTION Franklin McLean Memorial Research Institute
ADDRESS 5841 S. Maryland Avenue, Chicago, Illinois 60637
TEL TELEX
IN CHARGE S. J. Gatley REPORTED BY A. J. Creer/N. Odeh

HISTORY AND STATUS

DESIGN, date 1965 Model tests 1967
ENG DESIGN, date 1965-67
CONSTRUCTION, date 1969
FIRST BEAM, date (or goal) July, 1969
MAJOR ALTERATIONS Deflector

COST, ACCELERATOR 240,000
COST, FACILITY, total 600K
FUNDED BY Department of Energy

ACCELERATOR STAFF, OPERATION AND DEVELOPMENT

SCIENTISTS 5 ENGINEERS 11
TECHNICIANS 4 CRAFTS 1
GRAD STUDENTS involved during year -

OPERATED BY X Research staff or Operators
OPERATION 15 hr/wk, On target 10 hr/wk
TIME DISTR. in house 100 % outside - %

BUDGET, op & dev
FUNDED BY
RESEARCH STAFF, not included above

USERS, in house Yes outside -
GRAD STUDENTS involved during year 4
RESEARCH BUDGET, in house

FUNDED BY

MAGNET

POLE FACE, diameter (compact) 81 cm, R-extraction 35 cm
R injection cm

GAP, min 5 cm, Field 20 kG }
max 10 cm, Field 12 kG } at 2×10^5
AVERAGE FIELD at R ext 16 kG Ampere turns

B max / < B > 1.25

NUMBER OF SECTORS [compact separated] Spiral, max deg
SECTOR ANGLE (ISS) deg

TRIMMING COILS 3 ea. 8 Turns 100A max.

CONDUCTOR, material and type AL Foil 1 mm

STORED ENERGY (cryogenic) MJ -4

POWER: main coils 58 max kW; current stability 5 x 10
trimming coils max kW; current stability

WEIGHT: Fe 14 tons; coils tons

COOLING system Water

ION ENERGY (Bending limit) E/A = q^2/A^2 MeV/amu
(Focusing limit) E/A = q/A MeV/amu

ACCELERATION SYSTEM

DEES, number 2 angle deg

BEAM APERTURE 2 cm, DC Bias 1.5 KV kV

TUNED by, coarse MS line VC Trimmer -4

RF 12 to 25 MHz, stable $\pm 1 \times 10$

Orb F 12 to 25 MHz

HARMONICS, RF/Orb F, used

DEE-Gnd, max 30 kV, min gap cm

STABILITY, (pk-pk noise)/(pk RF volt)

ENERGY GAIN, max 60 max kV/turn

RF PHASE, stable to \pm deg

RF POWER input, max 29 kW

FREQUENCY MODULATION, rate /s

modulator, type

beam pulse, width

VACUUM SYSTEM

OPERATING PRESSURE 1×10^{-5} Torr or mbar

PUMPS, No, Type, Size 1 ea. 10" Oil Diffusion,

1 ea. 21 CFM Mechanical

ION SOURCES

Ion. Heated Pig.

INJECTION SYSTEM

None

EXTRACTION SYSTEM

Electrostatic Channel with Compensated Iron Chann

FACILITIES FOR RESEARCH

SHIELDED AREA, fixed 62 m², movable - m²

TARGET STATIONS 2 in 2 rooms

STATIONS served at same time, max 1

MAG SPECTROGRAPH, type -

COMPUTER model -

OTHER FACILITIES

CHARACTERISTIC BEAMS

PARTICLE	ENERGY (MeV)		CURRENT (pA)	
	Goal	Achieved	Internal	External
Proton	15	14.8	110	55
Deut	8	8.3	400	270
He 3 ⁺⁺	20	20.3	120	53
He 4 ⁺⁺	15	15	80	40

SECONDARY (part/s)

BEAM PROPERTIES

MEASURED	CONDITIONS	
	RF deg	MeV
PULSE WIDTH	μ A of	ions
PHASE EXC, max	μ A of	ions
EXTRACT eff. 55 %	270 μ A of	8 MeV D ions
RESOL $\Delta E/E$.1 %	μ A of	MeV ions

EMITTANCE

50 axial (r mm-mrad) 50 rad 90 μ A of MeV

OPERATING PROGRAMS, time distribution

BASIC NUCLEAR PHYSICS SOLID STATES PHYSICS

BIOMEDICAL APPLICAT 100% ISOTOPE PRODUCTIONS

REFERENCES/NOTES In AIP Conference Proceedings, #9, 1

- 1) Compact Cyclotron Engg. G.O. Hendry
- 2) ACRH Cyclotron, P.V. Harper
- 3) Design of Neutron Therapy Facility, F.T. Kuchnier

PLAN VIEW OF FACILITY, COMMENTS, ETC.

1. ³He recovery system for economical ³He⁺⁺ operation
2. Particle changes are made in 30 minutes
3. Targets may be irradiated internally or externally
4. Two external target stations; one for isotope, the other for neutron production
5. External beams transport system includes two quadrupole doublets, one steering magnet, one switching magnet, and four collimators