

ENTRY NO. C39 Date September 1995
 Name of Machine MINI-CYCLOTRON ILEC
 Institution Eindhoven University of Technology (EUT)
 Address Cyclotron Laboratory, EUT, P.O. Box 513, 5600 MB Eindhoven, The Netherlands
 Tel (040) 2474048 Telex 51163 Fax 438060 E-MAIL
 In Charge: J.A. van der Heide Reported by: J.A. van der Heide

HISTORY

MILESTONE DATES:

Design 1982 Model Tests 1986
 Construction 1985 First Beam 1989

DESIGN/CONSTRUCTION BY:

in house yes other no
 COST: Accelerator DFL 400.000 Facility DFL 500.000
 FUNDED BY: University

STATUS

STAFF: Machine

Scientists 1 Engineers 1
 Technicians 1 Students 2
 Research (in house/external)

Scientists / Engineers /
 Technicians / Students /
 BUDGET: Machine DFL 20.000 Funded by EUT
 Research Funded by

TIME DISTRIBUTION:

Basic Research (in house/external) % / %
 Applied Program (in house/external) 25 % / %
 Maintenance 25 % Development 50 %

MAGNET

POLE PARAMETERS:

Diameter 20 cm R_{extract} 16.7 cm R_{inject} cm
 HILL PARAMETERS: Gap (min) 3.3 cm B_{max} 1.8 T
 (@ AT) Gap (max) cm B_{min} T

VALLEY PARAMETERS: Gap (min) 5.0 cm B_{max} 1.2 T
 (@ AT) Gap (max) cm B_{min} T

AVERAGE FIELD: $\langle B \rangle_{\text{min}}$ 1.43 T $\langle B \rangle_{\text{max}}$ 1.43 T

NUMBER OF SECTORS: compact/separated 4 /
 sector angle 40° hill deg spiral (max) 0 deg

FIELD TRIMMING: Trim Coils no circular
 Harmonic Coils 4 sets
 Other

CURRENT: Main Coils 140 Amps Stability 10⁻⁵
 Trim Coils 20 Amps Stability 10⁻⁴
 Stored Energy (cryogenic) MJ

WEIGHT: Iron 2.8 tons Conductor 0.25 ton

ION ENERGY: Bending Limit E/A = 3 q²/A² MeV/u
 Focusing Limit E/A = q/A MeV/u

ACCELERATION SYSTEM

FUNDAMENTAL ACCELERATION:

Description: 2nd harmonic dees
 No. of Gaps/turn 4 dE/dn(max) 0.080 MeV/q
 Voltage (max) 0.028 MV Harmonic $f_{\text{rf}}/f_{\text{ion}}$ 2
 Freq 43.5 MHz Power in(max) 0.014 MW
 Stability: Phase 0.2° Voltage 10⁻⁴

OTHER CAVITIES (Flattopping or otherwise):

Description: two sixth harmonic flattop dees
 Region of Influence: R_{min} 5.5 cm R_{max} 16.7 cm
 No. of Gaps/turn 4 dE/dn(max) 0.012 MeV/q
 Voltage (max) 0.003 MV Harmonic $f_{\text{rf}}/f_{\text{ion}}$ 6
 Freq 130.5 MHz Power in(max) 0.002 MW
 Stability: Phase 0.7° Voltage 10⁻⁴

VACUUM SYSTEM

OPERATING PRESSURE: 10⁻⁵ Torr

PUMPS: (No. and type) 1 oil diffusion pump
 3000 l.s⁻¹

ION SOURCE(S)

Type	Intensity (mA)	@ $\epsilon_n = \beta \gamma c$ (mm mrad)	Ion Species
(a) Internal	3		protons
(b)			
(c)			
(d)			

INJECTION SYSTEM

Efficiency %

EXTRACTION SYSTEM

Electrostatic and passive magnetic Efficiency 50 %

CHARACTERISTIC BEAMS

Accelerated Ions	E/A (MeV/u)	Current (part μ A)	
		Internal	External
(a) protons	2.9	60	30
(b)			
(c)			
(d)			

Secondary Particles	E (MeV)	part/sec
(a)		
(b)		
(c)		

EXTRACTED BEAM PROPERTIES:

For μ A of MeV/u ions
 $\Delta E/E$ % $\Delta \phi$ °
 $\epsilon_n = \beta \gamma c$ x mm mrad z mm mrad

FACILITIES FOR RESEARCH

SHIELDED AREA: Fixed: 30 m² Moveable 60 m²
 Target Stations: 3 No. Served At Same Time: 1

MAGNETIC SPECTROMETERS:

OTHER FACILITIES:

REFERENCES/NOTES

- J.A. van der Heide, e.a. Nucl. Instr. & Meth. A240 (1985) 32-35 and J.A. van der Heide e.a. Nucl. Instr. & Meth. B64 (1992) 336-341

PLAN VIEW OF FACILITY, COMMENTS

The pole faces are copper plated to form grounded parts of the RF circuit.

