

ENTRY NO. 058 Date June 1992  
 Name of Machine U-240 Cyclotron  
 Institution Institute for the nuclear Research Ukrainian Academy of Sciences  
 Address Ukraine, Kiev - 252028, prospekt Nauki, 47  
 Tel (044) 2651456 Telex 132400 Proton Fax (044) 2654463 EMAIL  
 In Charge: I.N. Vishnevsky Reported by: A.I. Papash

**HISTORY**  
 MILESTONE DATES:  
 Design 1965 - 1970 Model Tests 1963 - 1966  
 Construction 1966 - 1972 First Beam March 1976  
 DESIGN/CONSTRUCTION BY:  
 in house Yes other Efremov's Institute/Leningrad  
 COST: Accelerator Facility 50 MRb(\$)  
 FUNDED BY: USSR Academy of Sciences  
 Committee Utilization of Atomic Energy

**STATUS**  
 STAFF: Machine  
 Scientists 6 Engineers 20  
 Technicians 60 Students 2  
 Research (in house/external)  
 Scientists 200 / Engineers /  
 Technicians / Students /  
 BUDGET: Machine 2 MRub (1990) Funded by Ukr. Ac. of Scien  
 Research Funded by -ces  
 TIME DISTRIBUTION:  
 Basic Research (in house/external) 50 % / %  
 Applied Program (in house/external) 20 % / %  
 Development % Maintenance 30 %

**MAGNET**  
 POLE PARAMETERS:  
 Diameter 240 cm R<sub>extract</sub> 102 cm R<sub>inject</sub> 2.3 cm  
 HILL PARAMETERS: Gap (min) 23.2 cm B<sub>max</sub> T  
 (0.83e6 AT) Gap (max) cm B<sub>min</sub> T  
 VALLEY PARAMETERS: Gap (min) 53.2 cm B<sub>max</sub> 11.2 T  
 (0 AT) Gap (max) cm B<sub>min</sub> T  
 AVERAGE FIELD: < B ><sub>min</sub> T < B ><sub>max</sub> 1.7 T  
 NUMBER OF SECTORS: compact/separated 3 /  
 sector angle deg. spiral (max) 45 deg.  
 FIELD TRIMMING: Trim Coils 15 circulating coils  
 Harmonic Coils 3 per sector  
 Other valley coils (1 per sector)  
 CURRENT: Main Coils 2100 Amps Stability 2.0e-5  
 Trim Coils 2100 Amps Stability 1.0e-4  
 Stored Energy (cryogenic) MJ  
 WEIGHT: Iron 650 tons Conductor 84 tons  
 ION ENERGY: Bending Limit E/A = 140 q<sup>2</sup>/A<sup>2</sup> MeV/u  
 Focussing Limit E/A = 80 (prptous) q/A MeV/u

**ACCELERATION SYSTEM**  
 FUNDAMENTAL ACCELERATION:  
 Description: 180 deg. and dummy deg.  
 No. of Gaps/turn 2 dE/dn(max) 0.18 MeV/q  
 Voltage(max) 0.09 MV Harmonic f<sub>rf</sub>/f<sub>ion</sub> 1.3  
 Freq 7 - 21 MHz Power in(max) 0.25 MW  
 Stability: Phase ± 20 Voltage 2.e-3  
 OTHER CAVITIES (Flattopping or otherwise):  
 Description:  
 Region of Influence: R<sub>min</sub> cm R<sub>max</sub> cm  
 No. of Gaps/turn dE/dn(max) MeV/q  
 Voltage(max) MV Harmonic f<sub>rf</sub>/f<sub>ion</sub>  
 Freq MHz Power in(max) MW  
 Stability: Phase Voltage

**VACUUM SYSTEM**  
 OPERATING PRESSURE: 2.0 e-6  
 PUMPS: No. and type 3 diffusion pumps (50 cm)

**ION SOURCE(S)**  
 Type Intensity (mA) ε<sub>n</sub> = βγε (πmm mrad) Ion Species  
 (a) internal pig  
 (b) with filament 2 r, u, a  
 (c) with indirect C<sub>4</sub><sup>+</sup>, B<sub>4</sub><sup>+</sup>, N<sub>5</sub><sup>+</sup>, N<sub>4</sub><sup>+</sup>  
 (d) heated cathode 0.1-1.0 E<sub>4</sub><sup>+</sup>, O<sub>4</sub><sup>+</sup>, Ne<sub>4</sub><sup>+</sup>, etc.

**INJECTION SYSTEM**  
 Axial injection... mirror inflector Efficiency %  
 not in operation  
**EXTRACTION SYSTEM**  
 e/s. deflector, compensated magnetic Efficiency 20-50 %  
 channel and iron channel

**CHARACTERISTIC BEAMS**

Accelerated Ions	E/A (MeV/u)	Current(part μA)	
		Internal	External
(a) protons	20-78	up to 200	10
(b) D <sub>1</sub> <sup>+</sup>	10-25	10-20	5-10
(c) C <sub>4</sub> <sup>+</sup> , O <sub>4</sub> <sup>+</sup> , Ne <sub>4</sub> <sup>+</sup>	5	2	0.5
(d) N <sub>4</sub> <sup>+</sup>	10	1	0.3

Secondary Particles E (MeV) part/sec  
 (a)   
 (b)   
 (c)

**EXTRACTED BEAM PROPERTIES:**  
 For 5 μA of 70 MeV/u R ions  
 ΔE/E ± 0.3 % Δφ 30°  
 ε<sub>n</sub> = βγε x 5 πmm mrad z 5 πmm mrad

**FACILITIES FOR RESEARCH**  
 SHIELDED AREA: Fixed 1000 m<sup>2</sup> Moveable 1200 m<sup>2</sup>  
 Target Stations: 15 No. Served At Same Time: 1  
 MAGNETIC SPECTROMETERS: MS 200 R=200 cm dE/E=20e-5  
 OTHER FACILITIES:

**REFERENCES/NOTES**  
 (a) Atomnaja Energiya, 6, 1976  
 (b) 8. International Conference on Cyclotrons, Blooming ton, September 18-21, 1976

**PLAN VIEW OF FACILITY, COMMENTS**  
 1. The polarized ion source will be installed in 1993  
 2. The ECR ion source in construction ( colloboration with JINR, Dubna )  
 3. The radioactive ions seperator in design ( Efremov NIIIEPA - Petersburg )  
 4. The storage ring complex with U-240 injector under consideration.